

Highlighting innovative design features and useful application information for **Thermo Scientific CO₂ Incubators.**

smart notes

design and innovation ► CO₂ Incubation

Thermo
SCIENTIFIC



Q A

Why is automated heat decontamination a better solution for sterilizing my CO₂ incubator than an automated chemical disinfection, such as *in situ* hydrogen peroxide vapor?

A CO₂ incubator automated heat decontamination cycle delivers ease-of-use and proven effectiveness. After removing cells and any heat sensitive items from the incubator, simply push a button to begin the automated heat decontamination cycle. In contrast, *in situ* hydrogen peroxide (H₂O₂) requires manual handling and a continuing investment in reagents, yet the effectiveness is uncertain.

Hydrogen peroxide vapor is often used for biological safety cabinets and room disinfection because heat decontamination is not possible due to the scope and size of these areas. This method uses external equipment to generate a concentrated vapor, and, due to the toxic nature of the H₂O₂, should be performed by trained personnel.

The “automated” in-chamber hydrogen peroxide vapor available on some CO₂ incubators is not the same as external H₂O₂ generation. This *in situ* H₂O₂ disinfection technology requires the user to handle the chemical and set up the H₂O₂ generator manually. This also involves disassembly and careful repositioning of all internal components – as much manual labor as separately autoclaving all of these parts. This additional handling may result in error or insufficient disinfection and can easily reintroduce contamination to the incubator chamber, jeopardizing cultured cells.



Automated Heat Decontamination

Hydrogen peroxide poses dangers for humans and cultured cells

Hydrogen peroxide vapor is toxic to humans and cultured cells as well as to microorganisms. This method for disinfecting laboratory equipment should be performed by a trained technician. The American Cancer Society states, "Direct skin contact with food grade hydrogen peroxide can cause blistering or burns, and breathing its vapors can cause serious breathing problems up to 72 hours later." (www.cancer.org, "Oxygen Therapy") The U.S. Occupational Safety and Health Association (OSHA) has set a Permissible Exposure Limit (PEL) of 1 part per million (ppm) per day exposure. Hydrogen peroxide can generate dangerous free radicals, damaging DNA and causing chromosomal aberrations. Based on these effects, any residual hydrogen peroxide must be entirely eliminated to avoid any detriment to cultured cells.

Dangers of Hydrogen Peroxide (as summarized from OSHA)	
Routes of Exposure:	Reference
Inhalation of vapor or mist, contact with eyes or skin, or ingestion	Sittig M 1991. Handbook of toxic and hazardous chemicals. 3rd ed. Park Ridge, NJ: Noyes Publications.
Health Hazard Effects:	Reference
Eye, skin and mucous membrane irritant	NIOSH 1995. Registry of toxic effects of chemical substances: Hydrogen Peroxide. Cincinnati, OH: U.S. Department of Health and Human Services, Public Health Service, Centers for Disease Control.
Inhalation of 7 ppm causes lung irritation in humans; exposure to vapor or mist for even short periods can cause eye stinging and tearing.	NLM 1995. Hazardous substances data bank: Hydrogen Peroxide. Bethesda, MD: National Library of Medicine. Hathaway GJ et al. 1991. Proctor and Hughes' chemical hazards of the workplace. 3rd ed. New York, NY: Van Nostrand Reinhold.
Mutagenic; damages DNA and causes chromosomal aberrations in cultured mammalian cells.	IARC 1985. IARC monographs on the evaluation of carcinogenic risk of chemicals to man. Volume 36. Lyon, France: World Health Organization, International Agency for Research on Cancer.
Liquid H ₂ O ₂ in contact with skin can cause whiteness or bleaching. If not washed promptly, irritation, redness and blisters can develop.	Hathaway GJ, Proctor NH, Hughes JP, and Fischman ML [1991]. Proctor and Hughes' chemical hazards of the workplace. 3rd ed. New York, NY: Van Nostrand Reinhold.

Table 1: Hydrogen peroxide poses dangers to humans and to cultured cells, as indicated by data summarized here. Data from the U.S. Occupational Safety and Health Association at www.osha.gov.

Choose a method with independent test results for proven effectiveness

An automated system for sterilization of a CO₂ incubator should provide consistent, measurable results that prove successful elimination of contamination. The effectiveness of automated hydrogen peroxide disinfection can be questionable because H₂O₂ molecules would have to make contact with every microorganism to be successful. Any microscopic areas missed by the H₂O₂ allow microorganisms to survive and propagate again. In contrast, Thermo Scientific™ CO₂ incubators provide independent proof of their high temperature cycles using accepted test microorganisms including bacterial spores. Application notes documenting these independent test results can be found at thermoscientific.com/co2.

Summary

Do-it-yourself hydrogen peroxide disinfection requires as much manual handling as autoclaving parts from a basic incubator. Any remaining residues can pose risk to cultured cells and humans and must be removed. This method is not proven to effectively eliminate contamination with the same assurance or ease of use that heat decontamination offers.

See how Thermo Scientific CO₂ incubators provide optimal cell growth. Learn more at www.thermoscientific.com/co2incubators

© 2014 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries.

Australia +61 39757 4300
Austria +43 1 801 40 0
Belgium +32 53 73 42 41
China +800 810 5118 or
+400 650 5118
France +33 2 2803 2180
Germany national toll free 0800 1 536 376
Germany international +49 6184 90 6000

India toll free 1800 22 8374
India +91 22 6716 2200
Italy +32 02 95059 552
Japan +81 3 5826 1616
Netherlands +31 76 579 55 55
New Zealand +64 9 980 6700
Nordic/Baltic/CIS countries
+358 9 329 10200

Russia +7 812 703 42 15
Spain/Portugal +34 93 223 09 18
Switzerland +41 44 454 12 22
UK/Ireland +44 870 609 9203
USA/Canada +1 866 984 3766

Other Asian countries +852 2885 4613
Countries not listed +49 6184 90 6000

Thermo
SCIENTIFIC
Part of Thermo Fisher Scientific